



**INTERNATIONAL STANDARD ISO 26782:2009**  
**TECHNICAL CORRIGENDUM 1**

Published 2009-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Anaesthetic and respiratory equipment — Spirometers intended  
for the measurement of time forced expired volumes in humans**

**TECHNICAL CORRIGENDUM 1**

*Matériel d'anesthésie et de réanimation respiratoire — Spiromètres destinés au mesurage des volumes  
expiratoires forcés chronométrés chez les humains*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 26782:2009 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*.

---

*Page 13, Figure A.1*

In the key, replace point 2 " $t_0 = 0,75$  s" with " $t_0 = 0,075$  s".

**Pour plus d'infos, merci de nous contacter.**

**Association Sénégalaise de Normalisation**

4 Avenue Jean Jaurès, Immeuble El Hadj Omar DIA, 6 ème Etage  
Dakar - SENEGAL

**Tel: +221 33 829 58 25**

**Email: [asn@asn.sn](mailto:asn@asn.sn)**