



Methods for the calibration of vibration and shock transducers —

Part 12: Primary vibration calibration by the reciprocity method

TECHNICAL CORRIGENDUM 1

Méthodes pour l'étalonnage des transducteurs de vibrations et de chocs —

Partie 12: Étalonnage primaire de vibrations par méthode réciproque

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 16063-12:2002 was prepared by Technical Committee ISO/TC 108, *Mechanical vibration, shock and condition monitoring*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.

Page 17, Equation (B.21)

Insert “–” on the left hand side of Equation (B.21), so that it reads:

$$-F = \frac{u_1 y_{em} z_n}{y_m z_n + 1} \quad (\text{B.21})$$

Pour plus d'infos, merci de nous contacter.

Association Sénégalaise de Normalisation

4 Avenue Jean Jaurès, Immeuble El Hadj Omar DIA, 6 ème Etage
Dakar - SENEGAL

Tel: +221 33 829 58 25

Email: asn@asn.sn