



INTERNATIONAL STANDARD ISO 7393-1:1985
TECHNICAL CORRIGENDUM 1

Published 2001-03-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Water quality — Determination of free chlorine and total chlorine —
Part 1:
Titrimetric method using *N,N*-diethyl-1,4-phenylenediamine

TECHNICAL CORRIGENDUM 1

Qualité de l'eau — Dosage du chlore libre et du chlore total —

*Partie 1: Méthode titrimétrique à la *N,N*-diéthylphénylène-1,4 diamine*

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 7393-1:1985 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

Page 2

In **4.5.2 Standardization of the solution**, replace the existing formula with the following:

$$c_1 = V_2 \frac{c_2}{2V_1}$$

Pour plus d'infos, merci de nous contacter.

Association Sénégalaise de Normalisation

4 Avenue Jean Jaurès, Immeuble El Hadj Omar DIA, 6 ème Etage
Dakar - SENEGAL

Tel: +221 33 829 58 25

Email: asn@asn.sn